

# Lifeloc FC10<sup>Plus</sup>

## Operations Manual

Unlock the Power  
of Alcohol Testing

### Introduction

Congratulations.....	4
Front View.....	5
Features.....	6

### Preparation

Installing Batteries.....	7
Turning the FC On and Off.....	7
Observing the Subject.....	8
Attaching a Mouthpiece.....	8

### Breath Testing Modes

Testing Modes Explained.....	9
End of Breath & Precise Volume Explained.....	9
Conducting an Automatic Test.....	10
Manual Override During an Automatic Test.....	10
Conducting a Manual Test.....	11
Conducting a Passive Test.....	11

### Operation

Breath Flow.....	12
Alcohol Curve.....	12
Test Results.....	12
Viewing Previous Test Results.....	12
Removing the Mouthpiece.....	12

### User Settings

Changing the Time.....	13
Changing the Date.....	13
Setting the Trigger Mode.....	14
Security Settings Explained.....	14
Setting Up Security Password.....	14
Using Security Password.....	15
Removing Security Password.....	15
Adjusting LCD Contrast.....	16
Setting the Default Test Order.....	16
Results Format Explained.....	17
Setting Results Format.....	17
Setting the Pass/Warn/Fail Levels.....	18

### Calibration Settings

Wet Bath and Dry Gas Explained.....	18
Selecting the Calibration Standard.....	19

Calibration/Cal Check Explained.....	19 - 20
Selecting the Standard Type.....	20
Calibration/Cal Check Lock-out Explained .....	20
Setting the Calibration Lock-out Type .....	21
Setting the Cal Check Lock-out Type.....	21
<b>Wet Bath Calibration</b>	
Wet Bath Simulator Set-Up .....	22
Wet Calibration.....	22 - 23
Wet Calibration Check.....	23 - 24
<b>Dry Gas Calibration</b>	
Dry Gas Tank Set-Up.....	24
Dry Gas Calibration.....	24 - 25
Dry Gas Calibration Check.....	26
<b>Computer Connection</b>	
AlcoMark <sup>®</sup> Explained .....	27
AlcoMark <sup>®</sup> Set-Up .....	27
<b>Maintenance</b>	
How to Check Status.....	27
Fuel Cells.....	28
Cleaning .....	28
Batteries .....	28
<b>On-Screen Messages</b>	
Messages Explained .....	29
<b>Service</b>	
Lifeloc Factory Warranty.....	30 - 32
Extended Service Plans.....	33
Service .....	33
<b>General Usage</b>	
Notice.....	33
Specifications.....	34
<b>FC10<sup>Plus</sup> Supplies</b>	
Accessories and Supplies .....	34
<b>Menu Charts</b>	
User Settings Chart.....	35
Calibration Display Chart.....	36

## Attention FC10<sup>Plus</sup> Operator:

---

### **Congratulations on your purchase of a Lifeloc FC10<sup>Plus</sup>.**

For over 30 years, Lifeloc Technologies has been providing advanced alcohol testing equipment & training to Law Enforcement and Corrections Professionals. We are the leader in product innovation, precision instruments, ease of use & Five Star Customer Care.

The FC10<sup>Plus</sup> breath alcohol tester is manufactured in Wheat Ridge, Colorado, by Lifeloc Technologies, Inc. Lifeloc offers premium quality products combined with exceptional service and technical support.


The FC10<sup>Plus</sup> is a state-of-the-art breath alcohol tester that is software based and incorporates unique cutting edge technologies. Because of the advanced FC10<sup>Plus</sup> design:

- Results on a positive test register in 10 seconds.
- You can take another test 30 seconds after a positive. Repeatability is not compromised.
- Your FC10<sup>Plus</sup> will automatically take the test when it senses a deep lung sample is delivered.
- Your FC10<sup>Plus</sup> will provide an accurate test, or else explain to you why it cannot, and even provide suggestions on how to proceed to complete an accurate test on your subject.
- AA Alkaline or NiMH batteries last for about 160 “on” hours or up to 6000 tests.

The following pages will explain in detail the operation of your FC10<sup>Plus</sup> Series portable breath tester.

## Unlock the Power of Alcohol Testing

## Front View (with Mouthpiece)



## Features

---

- **Large Graphic LCD Display:** Capable of showing numbers, letters, icons and plain English text messages.
- **Simplified Calibration:** Automatic with EASYCAL<sup>®</sup>. Software controlled guiding you step by step manually.
- **Auto Test Mode:** The easiest way to take a test. Tester automatically takes deep lung sample when subject is at end of breath.
- **Manual Test Mode:** Enables operator to control exact point of breath sample.
- **Passive Test Mode:** Checks for the presence of alcohol in the breath or in an open container, without using a mouthpiece.
- **On Board Memory:** Stores the last 100 tests.
- **Real Time Clock:** Stores time and date information with test results as well as calibration and cal check results.
- **User-Selectable Test Order:** Allows choice of either Auto Test or Passive Test default mode.
- **Calibration Reminder with Lockout:** Prevents you from using an FC10<sup>Plus</sup> when it is due for calibration.
- **Auto Shut-Off:** Preserves battery life.
- **Fast, Simple Operation:** While the FC10<sup>Plus</sup> contains a host of features, it is still easy to use.
- **Automatic Backlight:** Easy viewing of test results either day or night.
- **Exceptional Battery Life:** Up to 160 hours or 6000 test operation using four AA Alkline or NiMH rechargeable batteries.

## Installing Batteries

---

Press in and down on the battery door located on the back of the FC10<sup>Plus</sup>.


Install the four AA Alkaline batteries in the direction of the symbols in the battery case.

Close the case by pushing up on the battery door until it locks shut.

See page 28 for directions on using rechargeable batteries.

## Turning the FC10<sup>Plus</sup> On and Off

---



Press and hold the Power button on the bottom of the front of the unit until it beeps. The FC10<sup>Plus</sup> performs an automatic internal diagnostics check.

To turn the unit off, press and hold the **Power** button until it beeps twice. The unit will shut down.

**Note:** Momentarily pressing the Power button, when the unit is on, will return you to the main menu.



## Observing the Subject

---

The FC10<sup>Plus</sup> provides a highly accurate reading of breath alcohol acquired by sampling deep lung air. Readings will also detect residual mouth alcohol.

To prevent mouth alcohol from affecting a test, make certain that the subject is not allowed to put anything in their mouth for 15 minutes prior to taking a test.

If the subject just took a drink, a 15 minute observation period in which they are not permitted to put anything in their mouth should be observed before testing. This will ensure all residual alcohol from any source is completely dissipated and test results will be valid.

## Attaching a Mouthpiece


---

Remove the mouthpiece from its wrapper, making sure not to touch the end which the subject will be blowing into.

Attach the mouthpiece to the port on the back of the FC10<sup>Plus</sup>.

Line up the mouthpiece port over the holes in the back of the mouthpiece. Press in place.

Ensure it is securely attached.





### Breath Testing Modes Explained

---

The FC10<sup>Plus</sup> is capable of conducting Automatic, Manual, and Passive tests.

- **Automatic Test** is the easiest way to achieve a deep lung sample. The FC10<sup>Plus</sup> monitors the subject's breath and automatically takes the sample near the end of the breath flow.
- **Manual Test** is normally used only when the subject is unable to provide a sufficient breath sample for the automatic test.
- **Passive Test** is a quick screen to detect alcohol but is not designed to quantify the results. Passive results are reported as "POS" if alcohol is detected, "NEG" if alcohol is not detected. In this mode, no mouthpiece is required.

### End of Breath and Precise Volume Explained

---

In Auto Test mode, the unit can be set to take a sample either:

- When the subject nears the end of the exhalation (**End of Breath**)
- or —
- When it detects 1.5L of breath (**Precise Volume**)

Both will give accurate results. Precise Volume may work better with uncooperative subjects.

(To select "END OF BREATH" or "PRECISE VOLUME" mode, see Trigger Mode instructions on p.14)

### Conducting an Automatic Test

---

Turn the FC10<sup>Plus</sup> on.

Verify the display reads “AUTO TEST”.


Attach the mouthpiece to the back of the unit.

Instruct the subject to blow into the mouthpiece **firmly and steadily for as long as they can**. (But not necessarily as hard as they can.) The unit calculates volume and will give an error message if the subject cannot reach 1.3 liters.

Read the result.

After taking a test, the FC10<sup>Plus</sup> will display the results in large numbers on the display.

Press the **Function** button to return to the test mode.



### Manual Override during an Automatic Test

---

**Note:** This feature allows the completion of a test in the occasional instance when the subject may have diminished lung capacity and cannot activate the Auto Test.

Turn the FC10<sup>Plus</sup> on.

Attach a mouthpiece to the back of the unit and verify the display reads “AUTO TEST”.

Instruct the subject to blow into the mouthpiece **firmly and steadily for as long as they can**.

When they are near the end of their breath, press the **Execute** button.

Read the result.

### Conducting a Manual Test

---

Turn the FC10<sup>Plus</sup> on.

Attach the mouthpiece to the back of the unit.


Press the **Function** button until the display reads “MANUAL TEST”.

Instruct the subject to blow into the mouthpiece **firmly and steadily for as long as they can**.

When they are near the end of their breath, press the **Execute** button.

Read the result.

**Please note**, manual test mode is pressure activated. If the subject does not blow air into the mouthpiece, the test can not be conducted.



### Conducting a Passive Test (No Mouthpiece)

---

Turn the FC10<sup>Plus</sup> on.

Press the **Function** button, if necessary, until the display reads “PASSIVE TEST.”


Hold the FC10<sup>Plus</sup> sample port (orange-colored opening labeled “Port” on the back of the FC10) about 4 inches from the subject’s mouth.

Have the subject blow toward the port.

Press the **Execute** button while the subject is blowing.


Read the result. The unit will only display “POS” or “NEG.” It will not display the actual numerical result.

**Note:** A passive test can also be done over an open container to detect the presence of alcohol.



## Breath Flow


---



As the subject blows into the mouthpiece, the FC10<sup>Plus</sup> will show a graph of the breath flow on the display, as well as showing the amount of liters. Liters are shown in the upper left corner.

## Alcohol Curve


---



If the FC10<sup>Plus</sup> detects alcohol, the alcohol level is graphed and will be displayed before the result.

## Test Results

---




After the alcohol is graphed, the test result is displayed.

The result will remain on the screen until the **Function** or **Power** button is pressed.

The last 100 test results are retained in memory.

## Viewing Previous Test Results

---



Press the **Function** button until "LAST TEST RESULT" is displayed. The results of the last test will be displayed.

To view any of the last 10 tests use the **+ or -** buttons to scroll through.

Momentarily press the **Power** button to return to the testing mode.

## Removing the Mouthpiece

---

Remove the mouthpiece by pushing straight down on the tab at the right of the display screen. DO NOT BEND the tab.

## Changing the Time

---

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button.

Display reads “TIME” with the hour digits flashing.

Press the **+ or -** button to change the hour.

Press the **Execute** button to save your changes and move to minutes.

Press the **+ or -** button to change the minutes.

Press the **Function** button to save your changes.

Momentarily press **Power** to return to the testing mode.

## Changing the Date

---

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “DATE” with the month digits flashing.

Press the **+ or -** button to change the month.

Press the **Execute** button to save the month and move to the day.


Press the **+ or -** button to change the day.

Press the **Execute** button to save the day and move to the year.

Press the **+ or -** button to change the year.

Press the **Function** button to save your changes.

Momentarily press **Power** to return to the testing mode.



## Setting Trigger Mode

---

From the “SETTINGS” display, press the **Execute** button.

Press the **Function** button until the display reads “TRIGGER MODE”.

Press the **Execute** button to select between “END OF BREATH” and “PRECISE VOLUME”.

Press the **Function** button to save the setting. Momentarily press **Power** to return to the testing mode.

(User settings flow chart is shown on p. 36)

## Security Settings Explained

---

Password can be set to protect access to calibration, calibration check, calibration settings and user settings.

Without a password, users are still able to conduct Automatic, Manual and Passive tests; print test results; and check battery status and temperature.

## Setting Up Security Password

---

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “SECURITY SETTINGS”.

Press the **Execute** button. The unit is now ready to accept a password. Use the **+ or -** buttons to enter a number or letter.

Use the **Execute** button to move to the next digit. Password is limited to 8 characters.

Press the **Function** button to save your password. Momentarily press **Power** to return to the testing mode.

**Note: Record and store your password in a safe place. Lifeloc does not have access to your password. If password is lost, the only way to reset the unit is to call Lifeloc Tech Support at 720.317.2190.**

## Using Security Password

---

For access to Calibration, Calibration Check, Calibration Settings or User Settings.

Press the **Function** button until the display reads “CALIBRATION” or “SETTINGS”. Select the one you would like to access. You are limited to one at a time.

Press the **Execute** button. Display reads “SECURITY CODE”.

Enter your password using the **+ or -** buttons. Use the **Execute** button to move to the next digit.

Press the **Function** button to gain access. You now have access to all menus and settings.

## Removing Security Password

---

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “SECURITY CODE”.

Enter your password using the **+ or -** buttons.

Press the **Function** button until the display reads “SECURITY SETTING”.

Press the **Execute** button. Display reads “SECURITY SETTING” and shows your password in the lower left corner.

Clear your password using the **+ or -** buttons to change the letters or numbers to blank spaces. Use the **Execute** button to move to the next digit.

Press the **Function** button to clear the password.

Momentarily press **Power** to return to the testing mode.



## Adjusting LCD Contrast

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “DISPLAY SETTINGS”.

Press the **Execute** button. Display reads “LCD CONTRAST”.

Press the **+ or -** buttons to adjust the contrast of the text.

(0 = lightest; 10 = darkest)

Press the **Function** button to save your adjustment. Momentarily press **Power** to return to testing mode.

## Setting the Default Test Order

Test Order 1 (Default)	Test Order 2	Test Order 3
Auto Test Manual Test Passive Test	Passive Test Auto Test Manual Test	Auto Test Manual Test

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “DISPLAY SETTINGS”.

Press the **Execute** button. Display reads “LCD CONTRAST”.

Press the **Function** button. Display reads “TEST ORDER”.

Press the **Execute** button to toggle between Test Order 1, 2, and 3 above.

Press the **Function** button to save the changes.

Momentarily press **Power** to return to the testing mode.

**Please note:** By selecting Test Order 3 you will disable passive testing.

(User settings flow chart is shown on p. 36)

## Results Format Explained

Results for Auto and Manual tests can be displayed in a “NUMERIC” or “PASS/WARN/FAIL” (PWF) format. Passive test results, however, can only be displayed as “POS” & “NEG”.

**NUMERIC** results are in a 3-digit BAC format.

**PASS/WARN/FAIL** results require setting specific levels for Pass and Fail.


For example, if Pass is set at .040 and Fail is set at .080 (default settings on your FC10<sup>Plus</sup>), then:

Pass = .000 - .040

Warn = .041 - .079

Fail = .080 and above

*Sample settings only. See page 18 for instructions on setting the Pass/Warn/Fail levels.*



## Setting Results Format

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “DISPLAY SETTINGS”.

Press the **Execute** button. Display reads “LCD CONTRAST”.

Press the **Function** button until the display reads “RESULTS FORMAT”.

Press **Execute** to toggle between “NUMERIC” and “PWF”.

Press the **Function** button to save your setting.

Momentarily press **Power** to return to the testing mode.

## Setting the Pass/Warn/Fail Levels

---

Press the **Function** button until the display reads “SETTINGS”.

Press the **Execute** button. Display reads “TIME”.

Press the **Function** button until the display reads “DISPLAY SETTINGS”.

Press the **Execute** button. Display reads “LCD CONTRAST”.

Press the **Function** button until the display reads “PASS LEVEL”.

Press the **+ or -** button to set the BAC Pass level.

Press the **Function** button. Display reads “FAIL LEVEL”.

Press the **+ or -** button to set the BAC Fail level.

Press the **Function** button to save your settings.

Momentarily press **Power** to return to the testing mode.

**Note:** *An alcohol reading between the PASS and FAIL levels will read WARN.*

## Wet Bath and Dry Gas Explained

---

You can calibrate and check your FC10<sup>Plus</sup> using either the wet bath or dry gas method. You must first set your FC10<sup>Plus</sup> to recognize which method or which ‘Standard Type’ you will be using.

Once you choose the Standard Type, the FC10<sup>Plus</sup> will store that information in memory and you do not have to set it again unless you change to a different method of performing a calibration/calibration check.

Dry Gas Calibration requires that, prior to calibration, you enter the Corrected Standard Value based on your altitude or elevation. Using the chart on the outside of the canister, multiply the number next to your elevation by the standard.

Example:

- Denver, Colorado’s elevation is 5200 ft. above sea level and dry gas standard = .100 BAC.
- Correction factor from tank is .820
- Corrected Standard =  $.100 \times .820 = .082$

If you move to a significantly different altitude before calibrating, you will have to change the standard in the FC10<sup>Plus</sup>.

**Wet Bath Calibration** does not require altitude correction.

**Note:** The FC10<sup>Plus</sup> comes from the factory set for Wet Bath type and .100 BAC Calibration Standard.

## Selecting the Calibration Standard


---

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. The display reads either “WET CHECK” or ” DRY CHECK”.

Press the **Function** button until the display reads “CAL SETTINGS”.

Press the **Execute** button. The display reads “CAL STANDARD”.



Use the **+** or **-** button to change the number to the BAC level of standard you will be calibrating to. It should be the same as on the bottle of certified solution or as your altitude corrected standard, if using a dry gas tank. (See p. 18)

Press the **Function** button to save the settings. Momentarily press **Power** to return to the testing mode.

Once you set the standard, you do not have to set it again unless you change solutions or elevation (dry gas only).

## Calibration/Cal Check Explained

---

**Calibration of an FC10<sup>Plus</sup>** sets your unit to a known alcohol concentration to enable accurate BAC (Breath Alcohol Concentration) results.

You can use dry gas or wet bath solution to calibrate your FC10<sup>Plus</sup>. Most commonly used solutions are .100, .080 or .040 BAC.

The FC10<sup>Plus</sup> must be between 68° and 95° F (20° - 35° C) to calibrate.

Lifeloc recommends you calibrate your FC10<sup>Plus</sup>:

**Once** every 12 months, regardless of how many tests you have performed

**Or**, at intervals specified by your Internal Policies, Quality Assurance Plan, or State Regulations

**Or**, after two failed Calibration Checks.

**A Calibration Check** simply verifies the FC10<sup>Plus</sup> was calibrated correctly and is within the acceptable accuracy range. *Calibration check is also referred to as “External Calibration Check,” “Accuracy Check,” “Verification” and “Cal Check.”*

Lifeloc recommends you perform a calibration check on your FC10<sup>Plus</sup>:

**Once** every 30 days

**Or**, at intervals specified by your Internal Policies, Quality Assurance Plan, or State Regulations.

## Selecting the Standard Type

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. The display reads either “WET CHECK” or “DRY CHECK.”


Press the **Function** button until the display reads “CAL SETTINGS”.

Press the **Execute** button. The display reads “CAL STANDARD”.

Press the **Function** button until the display reads “STANDARD TYPE”.

Press the **Execute** button to choose between “DRY GAS” and “WET BATH”.

Press the **Function** button to save settings. Momentarily press **Power** to return to the testing mode.



## Calibration/Cal Check Lock-out Explained

The calibration/cal check lock-out allows you to prevent usage of your FC10<sup>Plus</sup> if it is not calibrated or cal checked in a specified time period. Time can be registered as days or number of tests.

Beginning 48 hours before the specified lock-out time, the unit will display “WARNING CAL EXPIRING,” “WARNING CHECK EXPIRING” or both.

When the time period has elapsed, the unit will display “CAL EXPIRED,” “CHECK EXPIRED” or both and not allow testing until it is calibrated and/or cal checked.

You can specify time periods up to 999 days or 9999 tests between calibrations or cal checks. You can also disable the calibration/cal check lock-out feature. (See p.21)

## Setting the Calibration Lock-out Type

---

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. Display reads either “WET CHECK” or “DRY CHECK”.

Press the **Function** button until the display reads “CAL SETTINGS”.

Press the **Execute** button. Display reads “CAL STANDARD”.

Press the **Function** button until the display reads “CAL TIME TYPE”.

Press the **Execute** button to select “DAYS” or “TESTS”.

Press the **Function** button to save your setting.

Display will now read “CAL TIME”, or “CAL NUM TESTS”.

Press the **+ or -** button to set the number of days or tests between calibrations. The time starts counting from your last calibration date, not from the date you set it. Choosing “DISABLED” turns this setting off.

Press the **Function** button to save your setting.

Momentarily press **Power** to return to the testing mode.

## Setting the Cal Check Lock-out Type

---

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. Display reads either “WET CHECK” or “DRY CHECK”.

Press the **Function** button until the display reads “CAL SETTINGS”.

Press the **Execute** button. Display reads “CAL STANDARD”.

Press the **Function** button until the display reads “CHECK TIME”, or “CHECK NUM TESTS”.

Press the **+ or -** button to set the number of days between calibration checks. The days start counting from your last calibration check date, not from the date you set it.

Press the **Function** button to save your settings.

Momentarily press **Power** to return to the testing mode.

### Wet Bath Simulator Set-Up For Calibration & Calibration Check

Pour a bottle of certified alcohol solution into the simulator jar and hand tighten lid.

Connect long tube from input port to output port so no alcohol escapes while simulator heats up.

Plug in the simulator and turn it on.

The simulator automatically heats the solution to 34° C (93.2° F) in about 5 to 10 minutes.

*Proper operating temperature is important for accuracy so be certain to check temperature before proceeding.*

### Performing a Wet Bath Calibration

Prepare the wet bath simulator according to its instructions above.

Disconnect long tube from the output port. Attach a mouthpiece adapter securely to the output port on the simulator lid. Refer to your simulator manual for location.

Turn the FC10<sup>Plus</sup> on. Attach a mouthpiece to the back of the unit.

Press the **Function** button until the display reads “CALIBRATION”, then press the **Execute** button. Display reads “WET CHECK”.

Press the **Function** button until the display reads “WET CALIBRATE”.

Verify the “CAL STANDARD” is set to the concentration of certified alcohol solution you will be using when you calibrate. To select a new Calibration Standard, see page 19.


Slide the FC10<sup>Plus</sup> mouthpiece over the mouthpiece adaptor on the simulator.

Be prepared to blow into the tube for up to 10 seconds.

#### Start blowing

- Blow through the input tube (or use a calibration pump) to create and maintain 1/2" of bubbles on the surface of the solution.
- Press **Execute** to proceed and start a 3 second countdown.
- Press **Execute** to take a sample.
- Continue blowing for another 3 seconds.

#### Stop blowing





If successful, the display will read “CAL COMPLETE”.

If no alcohol is detected, display reads “INVALID CALIBRATION”. Please repeat calibration setup and test, starting with wet bath simulator setup.

Disconnect the unit and wait at least two minutes before conducting a Cal Check to verify the accuracy of your calibration.

**Note:** Following calibration, a Cal Check is required before the unit allows you to conduct any tests.

(Calibration display flow chart is shown on p. 37)

## Performing a Wet Calibration Check

---

Prepare the wet bath simulator according to its instructions on page 22.

Disconnect long tube from the output port. Attach a mouthpiece adapter securely to the output port on the simulator lid. Refer to your simulator manual for location.

Turn the FC10<sup>Plus</sup> on. Attach a mouthpiece to the back of the unit.

Press the **Function** button until the display reads “CALIBRATION”, then press the **Execute** button. Display reads “WET CHECK”.

Verify the “CAL STANDARD” is set to the concentration of certified alcohol solution you will be using when you check the calibration. To select a new Calibration Standard, see page 19.

Slide the FC10<sup>Plus</sup> mouthpiece over the mouthpiece adaptor on the simulator.

Be prepared to blow into the tube for up to 10 seconds.

### Start blowing

- Blow through the input tube (or use a calibration pump) to create and maintain 1/2" of bubbles on the surface of the solution.
- Press **Execute** to proceed and start a 3 second countdown.
- Press **Execute** to take a sample.
- Continue blowing for another 3 seconds.

### Stop blowing

Read the result. It should be within +/- .005 BAC of the standard used. If your solution is greater than .100 BAC, accurate results will be within +/- 5%.

Example:

- A .100 BAC solution should read between .095 and .105 BAC.
- A .200 BAC solution should read between .190 and .210 BAC.

If no alcohol was detected, the display will read “INVALID CHECK”. Repeat wet bath calibration check instructions.

(Calibration display flow chart is shown on p. 37)


## Dry Gas Tank Set-Up

Attach the regulator to the dry gas tank.

Attach the small tube to the output port on the regulator.

Securely fit the mouthpiece adaptor to the small tube on the regulator.

**Note:** Do not store the dry gas tank with the regulator attached. The regulator is not designed as the primary sealing mechanism. Leaving the regulator attached may result in a leak.



## Performing a Dry Gas Calibration

Prepare the dry gas tank according to its instructions above.

Turn FC10<sup>Plus</sup> on. Attach a mouthpiece to the back of the unit.

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. Display reads “DRY CHECK”.

Press the **Function** button until the display reads “DRY CALIBRATE”.

Verify the Calibration Standard is set to the corrected BAC.

Using the Altitude Correction Factor (ACF) chart on the dry gas tank, calculate the corrected BAC.

Corrected Standard = (tank BAC) x (ACF)

To select a new Calibration Standard, see page 19.

Attach the FC10<sup>Plus</sup> mouthpiece to the regulator by sliding it over the mouthpiece adaptor, ensuring a snug fit.

Be prepared to press the **Gas Tank Regulator** button for up to 10 seconds.

**Start pressing** (gas flow will start)

- Press the **Execute** button to proceed and start a 3 second countdown.
- Press the **Execute** button again to take a sample.
- Continue holding the **Regulator** button for another 3 seconds.

**Stop pressing** (gas flow will stop)

If successful, the display will read “CAL COMPLETE”.

If no alcohol is detected, display reads “INVALID CALIBRATION”. Please check gas level on your tank regulator and replace tank if empty. Repeat dry gas calibration instructions on page 24.

Disconnect the FC10<sup>Plus</sup> and wait at least two minutes before conducting a Cal Check.

**Note:** Following calibration, a Cal Check is required before the unit allows you to conduct any tests.

(Calibration display flow chart is shown on p. 37)

## Performing a Dry Gas Calibration Check

---

Prepare the dry gas tank according to its instructions on page 24.

Turn the FC10<sup>Plus</sup> on. Attach a mouthpiece to the back of the unit.

Press the **Function** button until the display reads “CALIBRATION”.

Press the **Execute** button. Display reads “DRY CHECK”.

Verify the Calibration Standard is set to the corrected BAC.

Using the Altitude Correction Factor (ACF) chart on the dry gas tank, calculate the corrected BAC.

$$\text{Corrected BAC} = (\text{tank BAC}) \times (\text{ACF})$$

To select a new Calibration Standard, see page 19.

Attach the FC10<sup>Plus</sup> mouthpiece to the regulator by sliding it over the mouthpiece adapter, ensuring a snug fit.

Be prepared to press the **Gas Tank Regulator** button for up to 10 seconds.

**Start pressing** (gas flow will start)

- Press the **Execute** button to proceed and start a 3 second countdown.
- Press the **Execute** button again to take a sample.
- Continue holding the **Regulator** button for another 3 seconds.

Stop pressing (gas flow will stop)

Read the result. It should be within +/- .005 BAC of the corrected standard used. If your solution is greater than .100, accurate results will be within +/- 5%.

Example: A .082 BAC corrected standard should read between .077 and .087 BAC.

If no alcohol is detected, display reads “INVALID CHECK”. Please check gas level on your tank regulator and replace tank if empty. Repeat dry gas calibration check instructions.

(Calibration display flow chart is shown on p. 37)

### AlcoMark<sup>®</sup> Explained

---

AlcoMark is a software program developed by Lifeloc, exclusively for use with Lifeloc's breath testing equipment. AlcoMark software enables you to download, track, store and print test results from your FC10<sup>Plus</sup> to your computer. The AlcoMark CD comes with a cable and instruction manual.

**Data Download** - With AlcoMark, you can view and download any or all of the tests and any or all of the calibration events stored in the FC10<sup>Plus</sup> to a Microsoft Excel spreadsheet, a comma delimited text file or an XML file.

**Printout** - With AlcoMark, you can print any test result or calibration event stored in your FC10<sup>Plus</sup> through your computer's printer.

**Remote Diagnostics** - With AlcoMark, you can save time and money by having remote diagnostics performed after downloading information from your FC10<sup>Plus</sup> to your computer and e-mailing it to Lifeloc.

### AlcoMark<sup>®</sup> Set-Up

---

Install the software on your computer using the provided AlcoMark disc. Connect the FC10<sup>Plus</sup> to the computer using the provide serial to USB cable. You may now begin using AlcoMark.

### How to Check Status

---

This function allows you to check the status and internal conditions of your FC10<sup>Plus</sup>.

Press the **Function** button until display reads "STATUS".

Press the **Execute** button. Display shows Model, Serial Number, Software Version, and Software Release Date.

## Fuel Cells

---

Fuel cells are highly durable sensors that are capable of providing accurate breath alcohol results for years. There are, however, a few precautions you should take to make certain that these devices perform for the longest period of time possible.

**Use the device.** Fuel cells like moisture, so it is a good idea to take tests periodically to provide needed moisture to the fuel cell, especially in dry climates. You do not need alcohol, just breath.

**Avoid cigarette smoke!** Make certain no one is permitted to blow cigarette smoke into the unit. This can damage or destroy the fuel cell.

## Cleaning

---

Use of a mild disinfectant cleaner and a soft cloth on the outside of the case is recommended periodically to keep your unit clean. Do not use alcohol to clean the unit.

## Batteries

---

Your FC10<sup>Plus</sup> default setting is for four AA batteries. However, you do have the ability to use NiMH rechargeable batteries as well.

To use NiMH, insert the batteries as directed and power on the unit.

Press the **Function** button until the display reads "SETTINGS".

Press the **Execute** button. Display shows "TIME".

Press the **Function** button until the display shows "BATTERY TYPE".

Press the **Execute** button to change your battery settings.

Press the **Function** button to save your settings.

Momentarily press the **Power** button to return to the testing mode.

The four AA batteries in your FC10<sup>Plus</sup> should last for about 160 hours of "on" time or up to 6000 tests. It is recommended you use high-quality alkaline batteries with your unit.

### Messages Explained

Message	Explanation
<1.3L Retest or Try Manual Test	.....➤ Breath flow ended before the subject blew 1.3 liters of breath. Instruct subject to try again or use manual test mode.
>0.6	.....➤ BAC is unusually high and above 0.6 BAC. Subject may require medical attention.
Calibration/Cal Check Expired	.....➤ Calibrate or Cal Check the unit.
Calibration/Cal Check Expiring	.....➤ The unit is within 48 hours of the Calibration or Cal Check lockout time.
External Interference	.....➤ External interference has been detected. Move to a different location and try again. If error persists call Lifeloc Technical Support.
Flow Error - Retry & Blow Steadily	.....➤ Exhalation not complete or interrupted. Instruct subject to blow steadily as long as they can.
Invalid Calibration/Cal Check	.....➤ No alcohol was detected. Repeat setup, then retest.
Log Empty	.....➤ There are no results in the memory.
Low Battery	.....➤ Battery voltage is too low to take a test. Replace batteries.
Low Li Battery	.....➤ The internal clock battery is low. Contact Lifeloc Technical Support.
Pump Reset Needed	.....➤ The pump needs to be reset. Follow onscreen instructions.
Temperature	.....➤ When calibrating or taking a test, unit is outside of temperature limits. Calibration range = 68° - 95° F (20° - 35° C) Testing range = 32° - 130° F (0° - 55° C)
Timeout	.....➤ User blew for > 15 seconds. Try again, blow harder.



## Lifeloc Factory Warranty

---

The FC10<sup>Plus</sup> comes with a one year limited parts and labor warranty, effective on the date of purchase by the end-user.

### **The Warranty covers:**

- Parts and labor on covered repairs
- Software updates, as applicable
- Air freight back to the customer after the unit is repaired (U.S. only)

### **The Warranty does not cover:**

- Freight to the Lifeloc factory
- Misuse, abuse, negligence or accidents

Lifeloc Technologies, Inc. (“Lifeloc”) warrants to the buyer that at the time of shipment from Lifeloc’s facilities, all new Lifeloc Equipment purchased through Lifeloc or one of our authorized distributors will be free from defects in material and workmanship, under normal use and service, provided that the buyer gives Lifeloc written notice of any defect within twelve (12) months from original invoice date (the “Warranty”). The term “Lifeloc Equipment” includes all [portable hand-held and fixed station breathalyzers manufactured by Lifeloc], as well as all printers, keyboards, cables, cases, and power supplies purchased as part of a new Lifeloc kit configuration. Software is warranted to perform substantially in accordance with Lifeloc user manuals and to be free from defects in materials and workmanship under normal use and service for a period of twelve (12) months from original invoice date.

The Warranty does not apply if the product has been altered, customized, repaired, reported stolen or modified by someone other than a Lifeloc factory authorized technician, or if parts other than Lifeloc approved parts are used in replacement or repair.

The buyer’s exclusive remedy and Lifeloc’s sole liability for breach of the Warranty shall be repair, replacement or, at Lifeloc’s option, refund of the original purchase price paid by the buyer to Lifeloc for the Lifeloc Equipment that is shown, to Lifeloc’s reasonable satisfaction, to be defective in breach of the Warranty (“Defective” and, the defect causing the Lifeloc Equipment to be Defective, a “Defect”). Under no circumstance shall Lifeloc be liable for an amount that exceeds the lesser of the cost of replacement or the original purchase price paid by the buyer to Lifeloc for the Lifeloc Equipment. Lifeloc shall not be responsible for any customer software, customer settings or configuration data or customer test records resident in any products returned for service, repair, warranty or recertification.

Lifeloc assumes no risk for damage in transit. If Lifeloc determines that a Defect was outside of or not covered under the Warranty, Lifeloc will estimate repair and service charges and obtain the buyer's authorization prior to conducting any work to repair the Defective Lifeloc Equipment. Following repair, the Lifeloc Equipment will be returned to the buyer via standard ground transportation prepaid. Express charges, if authorized by the buyer, will be invoiced at the difference between the express charges and the standard return shipping charges.

**Lifeloc shall not incur costs related to loss, damage or incomplete or inaccurate paperwork of returned product regardless of origination point.**

**THIS WARRANTY IS THE PURCHASER'S SOLE AND EXCLUSIVE REMEDY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED. LIFELOC SHALL NOT BE LIABLE FOR ANY SPECIAL, INDIRECT, INCIDENTAL OR CONSEQUENTIAL DAMAGES OR LOSSES, INCLUDING LOSS OF DATA, WHETHER ARISING FROM BREACH OF WARRANTY OR BASED ON CONTRACT, TORT, RELIANCE OR ANY OTHER THEORY.**

NOTWITHSTANDING ANYTHING TO THE CONTRARY IN THE FOREGOING WARRANTY, OR IN ANY AGREEMENT BETWEEN THE BUYER AND LIFELOC PERTAINING TO ANY LIFELOC EQUIPMENT, OR AT LAW OR IN EQUITY:

THE WARRANTIES, OBLIGATIONS, AND LIABILITIES OF LIFELOC, AND THE REMEDIES OF THE BUYER SET OUT IN THE FOREGOING WARRANTY, ARE SOLE AND EXCLUSIVE, AND ARE MADE AND ACCEPTED BY THE BUYER IN LIEU OF, AND THE BUYER WAIVES AND RELEASES, ALL OTHER WARRANTIES, OBLIGATIONS, AND LIABILITIES OF LIFELOC, AND ALL OTHER CLAIMS AND REMEDIES OF THE BUYER, EXPRESS OR IMPLIED, ARISING BY STATUTE OR OTHERWISE, WITH RESPECT TO ANY DEFECT IN LIFELOC EQUIPMENT, INCLUDING BUT NOT LIMITED TO THE WAIVER AND RELEASE BY THE BUYER OF THE FOLLOWING: (i) ANY STATUTORY OR IMPLIED WARRANTY OF MERCHANTABILITY OR OF FITNESS FOR ANY PURPOSE, (ii) ANY IMPLIED WARRANTY ARISING FROM COURSE OF DEALING, COURSE OF PERFORMANCE OR USAGE OF TRADE, AND (iii) ANY OTHER RIGHT, CLAIM, OR REMEDY WHATSOEVER OF THE BUYER OR OF ANY PERSON OR ENTITY CLAIMING BY, THROUGH, OR UNDER THE BUYER AGAINST LIFELOC, WHETHER ARISING PURSUANT TO THE FOREGOING WARRANTY, PURSUANT TO ANY AGREEMENT PERTAINING TO LIFELOC EQUIPMENT, OR IN CONTRACT, IN INDEMNITY, IN TORT (INCLUDING, BUT NOT LIMITED TO, NEGLIGENCE), IN PRODUCTS LIABILITY, IN STRICT LIABILITY, OR OTHERWISE. BY ACCEPTING OR USING LIFELOC EQUIPMENT, THE BUYER FURTHER ACKNOWLEDGES, UNDERSTANDS, AND AGREES THAT NO STATEMENT OR REPRESENTATION HAS BEEN MADE BY LIFELOC, OR RELIED UPON BY THE BUYER, THAT IS INCONSISTENT WITH THE FOREGOING WARRANTY.

Without prejudice to the foregoing, Lifeloc shall not have any obligation or liability, and the buyer acknowledges that Lifeloc shall not have any obligation or liability whatsoever,

to the buyer, or to any person or entity claiming by, through, or under the buyer, whether arising pursuant to the foregoing Warranty, pursuant to any agreement pertaining to Lifeloc Equipment, or in contract, in indemnity, in tort (including, but not limited to, negligence), in products liability, in strict liability, or otherwise, (i) for any transportation, installation, removal, reinstallation, adjustment, or other expenses related to any Lifeloc Equipment covered by the Warranty or to other property, (ii) for any damage or loss to any property other than the Lifeloc Equipment covered by the Warranty, or (iii) for any special, indirect, incidental, or consequential damage or loss, even though such expenses, damages, or losses may be foreseeable, including, but not limited to: loss of profits or revenues, loss of use or equipment, cost of capital, cost of substitute equipment, repairs, or facilities, cost of downtime, or cost of purchased or replacement equipment or parts.

The limited remedies of the buyer set forth above shall be exclusive even though they may fail of their essential purpose. No agreement varying or extending the foregoing Warranty, no remedies, no exclusions, or no limitations shall be effective unless in a writing signed by an executive officer of LIFELOC. The correction of any Defect shall in no way extend the duration of the Warranty; only the unexpired warranty term of the Lifeloc Equipment applies to any repaired or replacement Lifeloc Equipment that is provided under the Warranty.

The Warranty is non-transferable and is effective on all Lifeloc Equipment purchased from and after [September 1, 2014]. Rights and recourse may vary by country.

## Extended Service Plans

---

Extended Service Plans are available for your FC10<sup>Plus</sup>. These provide complete coverage for an additional year at a reasonable cost and include free factory diagnostic checks. Call Lifeloc for details or purchase online at [www.lifeloc.com/esp](http://www.lifeloc.com/esp)

## Service

---

If your FC10<sup>Plus</sup> should require repairs or maintenance, Lifeloc is here for you. Just an email or phone call will put you in contact with our technical support personnel. Many minor adjustments can be made over the phone.

Repairs are completed within 5 days for up to 5 units or it's free.

## Notice

---


The FC10<sup>Plus</sup> is a professional device designed to be used by trained operators in conjunction with a specified, periodic maintenance and calibration/calibration check regimen. **Use by untrained operators or without periodic calibration or calibration checks may result in invalid results or incorrect interpretation of results. FC10<sup>Plus</sup> is not to be used by children under 12 years of age.**

**DO NOT DRINK AND DRIVE.** Lifeloc strongly recommends that no vehicle or machinery be operated after alcohol consumption. Even small quantities of alcohol can result in driving impairment.

The FC10<sup>Plus</sup> is not waterproof and should not be immersed in or exposed to excessive water. The FC10<sup>Plus</sup> is not suitable for use in a potentially explosive environment. The FC10<sup>Plus</sup> cannot be used inside an oxygen tent.

If instrument will not be used for more than 6 months remove battery to avoid damage to the instrument caused by leaking battery acid.

### Disposal of Instrument



At the end of the instrument's service life:

- Do Not dispose of the FC10<sup>Plus</sup> as unsorted municipal waste.
- Dispose of the FC10<sup>Plus</sup> in accordance with national waste disposal regulations.

## Specifications

Size.....	2.6" x 5" x 1.25" (66 x 127 x 32 mm)
Weight [w/ batteries] .....	9 oz. (255 grams)
Measurement Range .....	.000 to .600 BAC
Accuracy .....	±.005 BAC up to .100 BAC ±5% for .100 - .400 BAC
Battery Life .....	Approximately 160 hours or up to 6000 tests

We recommend storing the FC10<sup>Plus</sup> in temperatures between 32° - 130° F (0° - 55° C)

## Accessories & Supplies

Lifeloc offers a complete line of accessories and supplies for your FC10<sup>Plus</sup>, including:  
For the most current listing of supplies for your FC10<sup>Plus</sup>, go to:

- [www.lifeloc.com](http://www.lifeloc.com). Click on Order Products

You can place your order online or call Customer Service at 303-431-9500 or 800-722-4817.



Mouthpieces



Rubber Grip



Carrying Cases



Certified Simulator Solutions




Dry Gas Standards




EASYCAL<sup>®</sup> Calibration Station

## User Settings Chart



## Calibration Display Chart





12441 West 49th Avenue, Suite 4  
Wheat Ridge, CO 80033  
303.431.9500 | 800.722.4872  
fax: 303.431.1423  
[www.lifeloc.com](http://www.lifeloc.com)

**Hours of Operation: 7:30 am - 5:00 pm MST**

If you are calling outside of these hours, please leave us a voice message. We will contact you the following business day!

From our single location in Wheat Ridge, Colorado, we manufacture and service our products with you in mind. Every product is designed with quality and ease-of-use as our priorities. And our 5 Star Service is the fastest in the industry.

Lifeloc breath testers are used across the US and in over 35 countries. We carry all of the accessories and supplies for your breath test equipment. Please call us to reorder supplies and accessories or for information on purchasing additional testers.

# Lifeloc FC10<sup>Plus</sup>

## Operations Manual