

SAI Global hereby grants:

Lifeloc Technologies Inc.

12441 W. 49th Avenue, #4, Wheat Ridge, CO United States

StandardsMark Licence

Manufactured to:

AS 3547-1997 - Breath alcohol testing devices for personal use & AS 3547:2019 - Breath alcohol testing devices

"the StandardsMark Licensee" the right to use the STANDARDSMARK as shown below only in respect of the goods described and detailed in the Schedule which are produced by the Licensee or on behalf of the Licensee* and which comply with the appropriate Standard referred to above as from time to time amended. The Licence is granted subject to the rules governing the use of the STANDARDSMARK and the Terms and Conditions for certification and licence. The Licensee covenants to comply with all the Rules and Terms and Conditions.

Certificate No:SMKH22047

Issued: 15 February 2021

Expires: 29 June 2026

Originally Certified: 30 June 2011

Current Certification: 11 February 2021

Frank Camasta
Global Head of Technical Services
SAI Global Assurance





* For details of manufacture, refer to the licensee





SCHEDULE TO STANDARDSMARK LICENCE

SAI Global hereby grants:

Lifeloc Technologies Inc.

12441 W. 49th Avenue, #4, Wheat Ridge, CO United States

StandardsMark Licence

Manufactured to:

AS 3547-1997 - Breath alcohol testing devices for personal use

Model identification of the goods on which the STANDARDSMARK may be used:

Туре	Model Name	Calibration Period (Months)	Calibration Type	Date Endorsed
2	FC10	6 months	Service Agent - Wet bath or Dry gas	3 Jun 2016
2	FC10 Plus	6 months	Service Agent - Wet bath or Dry gas	3 Jun 2016
2	FC20	6 months	Service Agent - Wet bath or Dry gas	3 Jun 2016
2	FC20BT	6 months	Service Agent - Wet bath or Dry gas	3 Jun 2016
2	Hornet FC-5	6 months	Service Agent - Wet bath or Dry gas	13 Jan 2021
2	LifeGuard	6 months	Service Agent - Wet bath or Dry gas	3 Oct 2017
2	Lifeloc LT7	2 months	Service Agent - Wet bath or Dry gas	3 Dec 2020
2	Lifeloc LT7i	2 months	Service Agent - Wet bath or Dry gas	3 Dec 2020
2	Lifeloc LX9	2 months	Service Agent - Wet bath or Dry gas	3 Dec 2020

& AS 3547:2019 - Breath alcohol testing devices

Model identification of the goods on which the STANDARDSMARK may be used:

Туре	Model Name	Calibration Period (Months)	Calibration Type	Date Endorsed
2	LIFELOC LT7	6 months	Service Agent - Wet bath or Dry gas	13 Jan 2021
2	LIFELOC LT7i	6 months	Service Agent - Wet bath or Dry gas	13 Jan 2021

Certificate No: SMKH22047 Issued Date: 15 February 2021

This schedule supersedes all previously issued schedules

The STANDARDSMARK is a registered certification trademark of SAI Global Pty Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 680 George Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to www.saiglobal.com, for the list of product models.



^{*} For details of manufacture, refer to the licensee

SCHEDULE TO

STANDARDSMARK LICENCE

Туре	Model Name	Calibration Period (Months)	Calibration Type	Date Endorsed
2	LIFELOC LX9	6 months	Service Agent - Wet bath or Dry gas	13 Jan 2021

	End of Pecord	
	End of Record	

Certificate No: SMKH22047 Issued Date: 15 February 2021

This schedule supersedes all previously issued schedules

* For details of manufacture, refer to the licensee

The STANDARDSMARK is a registered certification trademark of SAI Global Pty Limited (A.C.N. 050 644 642) and is issued under licence by SAI Global Certification Services Pty Limited (ACN 108 716 669) ("SAI Global") 680 George Street, Sydney NSW 2000, GPO Box 5420 Sydney NSW 2001. This certificate remains the property of SAI Global and must be returned to SAI Global upon its request. Refer to www.saiglobal.com, for the list of product models.

